


Cleveland open cup flash point test



Complies with BS2000-36, EN ISO2592, ASTM D92 and IP36 standards.

The apparatus is applicable for determining measurement of flash points and burning points of petroleum products (the Cleveland open cup method) except petroleum products with flash points and burning points lower than 79°C

Specification

Model	SYD-3536	SYD-3536D
Heating apparatus	Quartz tube electric furnace heating 0-600W(Adjustable)	
Fuel	Gas	
Display	Dial plate	LCD
Nozzle aperture	0.6~0.8mm	
Temperature rang	-6~400°C	
Temperature accuracy	2°C	
Power	AC 220V 50Hz 650W	