

- Multi-functional strength tester is suitable for testing the concrete (C10~C80) strength and External wall tapestry brick bonding strength on site, and also suitable for testing the other pull strength. It adopts the Post-install pull-off test method for testing the concrete strength through testing the anchoring parts pull-off strength to calculate the concrete strength. Tapestry brick bonding strength is tested through three-point force of reaction to tapestry brick.

### STANDARD

EN 1542, EN 1348, EN 1015-12

### 2. Specification

Capacity	40KN	
Tripod inner diameter	120MM	
Accuracy	$\pm 1\%$ F.S	
Error		
Weight		
Test block	40*40*8mm	3pcs
Pull rod	1set	
Concrete anchor	1set	
Sander	1set	
Carry case	1pcs	



C S T INSTRUMENTS(THAILAND)LTD.,PART.

95 Romklao Road,  
Klongsam-Praves,  
Ladkrabang, Bangkok  
Tel.02-915-1693-5

Dealer