


**STANDARD: EN 1350-7, BS 1881-106, ASTM C231**

- This meter using press way to test the air contain in the testing sample of concrete. It is easy to read testing result of how many air contain in the testing sample of concrete. The meter not only can use to test the concrete, also can test the concrete with water filling in the meter. Without water then you can read the dial meter by red color scale, with filling water you can read black scales.

**• Specification**

Model	HC-7
Capacity	7L
Max. aggregate Size	40mm
Air content range	0-10%
Accurarcy	0-8% ( $\pm 0.1\%$ ), 8-10% ( $\pm 0.2\%$ )
Weight	10Kg

**C S T INSTRUMENTS(THAILAND)LTD.,PART.**

95 Romklao Road,  
Klongsam-Praves,  
Ladkrabang, Bangkok  
Tel.02-915-1693-5

**Dealer**